

Title;

Coupled LC-GC for the Analysis of Pesticide in Food

Ryoichi Sasano, Yutaka Nakanishi

Saika Technological Institute Foundation,

75-2, Kuroda, Wakayama-city, Japan

E-mail: sasano@saika.or.jp

Abstract;

The purpose of this study is to couple reversed-phase high performance liquid chromatography (LC) with gas chromatography (GC) for automated analysis of pesticide in food samples.

The interface employed for coupling LC with GC consists of two channels equipped with a solid phase extraction (SPE) cartridge. One is a channel where water is added to the eluate containing the target compound separated by LC and continuously it is loaded on the SPE cartridge. The target compound is adsorbed in the SPE cartridge. The other is channel where the adsorbed target compound is eluted from the SPE cartridge and the eluate is injected directly into the GC injector. By using SPE for coupling LC with GC, it becomes possible to change to the solvent which can be injected into the GC (e.g., acetone and hexane) from the solvent which cannot be injected (e.g., water and acetonitrile). The GC injector is a PTV injector equipped with a stomach type of insert in order to inject all of eluate from the SPE cartridge. This LC-(SPE)-GC system has been evaluated by analyzing chlorpyrifos in spinach. The recovery and repeatability were satisfactory.

Coupled LC-GC for the Analysis of Pesticide in Food

Ryoichi Sasano, Yutaka Nakanishi

Saika Technological Institute Foundation

75-2, Kuroda, Wakayama-city, Japan

Introduction

Analysis of pesticide residues in food is not so easy to do because various foods contain complex matrix. Recently several reports have been described that on-line LC-GC system is useful for the analysis of complex samples such as pesticide residues in food. The high sample capacity and wide range of separation performance of LC can be utilized in selective cleanup of sample and GC has high separation efficiency. However, coupling of LC to GC is not a trivial matter because the introduction of a large amount of LC fraction into a GC column requires the use of special techniques to separate the solvent from the sample. One of the subject for coupling LC with GC is a technique of eliminating water which is contained in LC fraction. Another subject is a technique of transferring analytes to GC.

The purpose of this study is coupling reversed-phase HPLC with GC for attaining automated analysis of pesticide residues in food. An interface equipped with a SPE cartridge is developed for coupling LC with GC. The system is shown in the following. With adding water to the LC fraction, the diluted fraction is loaded on the SPE cartridge. The analyte is adsorbed on the SPE. The cartridge was dried with nitrogen gas. The analyte is eluted from the the SPE cartridge with 20 μ L of hexane and the eluate is directly injected into a GC injector via a needle.

Finally, the performance of this LC-(SPE)-GC system has been evaluated by inspecting the data with analyzing chlorpyrifos in spinach.

Experimental


HPLC (MIDAS;Spark, Agilent 1100)
 Injection: 100 μ L, Sample loop
 Column: 2.1 mm i.d. \times 100 mm Inertsil ODS-3
 Solvents: A: Acetonitrile/water (50/50)
 B: Acetonitrile
 flow rate 0.5 mL/min
 Detector: UV 210 nm

Interface SPE
 SPE: 2 mm i.d. \times 10 mm C18
 Diluting: water 0.4 mL/min
 Purge: N2 gas, 1 min
 Elution: Hexane, 20 μ L

Interface Injector (LaviStoma; EMINET)
 Insert: Stomach Type Insert
 Solvent Vent: 10 sec, Purge flow 150 mL/min
 Splitless: 3 min
 Inj. Temp.: 70°C(3min)-120°C/min-220°C/min(3min)
 -50°C/min-260°C(10min)


GC/MS (QP-5050A; Shimadzu)
 Column: 0.25 mm i.d. \times 30 m Inert Cap 5MS/Sil
 Oven: 70°C(3min)-20°C/min-280°C(4min)
 Carr. gas: He, 1 mL/min
 MS: SCAN;50-550 m/z, SIM;199, 314 m/z

LC-(SPE)-GC system




SPE Interface

Dilution & Concentration




With adding water to the LC fraction, the diluted fraction is loaded on the SPE cartridge. The analyte is adsorbed on the SPE.

Elution & Injection



The analyte is eluted from the SPE with hexane and the eluate is directly injected into a GC injector via a needle.

PTV Interface



1st Stage Injection
 The injector is kept at a temperature lower than the boiling point of sample solvent.

2nd Stage Concentration
 While the evaporated sample solvent is exhausted, the sample is concentrated.

3rd Stage Transfer
 Target compounds are transferred to the column at an elevated temperature.

4th Stage Elimination
 Matrix compounds are eliminated from the insert at further elevated temperature.

Application

Sample preparation

20 g portions of homogenized spinach were extracted with 70 mL of acetonitrile. The extract solution was adjusted to 100 mL with water. Then 1 mL of the extract was cleaned up through an C18 cartridge (50 mg) for avoiding HPLC column from deteriorating, and adjusted to 2 mL with 70% acetonitrile-water for LC-(SPE)-GC/MS analysis.




Fig. 1 HPLC chromatogram of a spinach spiked with 1 μ g/g of chlorpyrifos (a) and a standard solution of it (b). Marked fraction transferred to the SPE interface.

Chlorpyrifos elute around 8.2min. The fraction eluting from 7.8 to 8.6 min was transferred to the SPE interface.





Fig. 3 MS spectrum of peak

Fig. 2 GC/MS-SCAN chromatogram of the LC-(SPE)-GC/MS analysis of a spinach spiked with 1 μ g/g of chlorpyrifos

The excellent chromatography obtained can be observed in Fig. 2. The cleaning achieved with the LC process is highly satisfactory. The performance of the system was investigated with respect to the LC-(SPE)-GC process by a standard solution or a spinach spiked with 1 μ g/g of chlorpyrifos. The recovery was found to be higher than 98%. Reproducibility parameters (RSD) for the peak area was below 2%.

Conclusion

Coupling of LC with GC has been accomplished by a new SPE interface and a PTV injector equipped with a stomach shaped insert. The LC-(SPE)-GC system provides very high efficiency and selectivity performance, and then allows automated analysis of pesticide residues in food.